INDIAN INSTITUTE OF TECHNOLOGY BOMBAY
MATERIALS MANAGEMENT DIVISION
Powai, Mumbai 400076.

Ref. PR No. 1000056408 Rfx. No. - 3600003072

Item Description – Repairing of Tap Reactor

Sr. No	Item Description	Detailed Technical Specification	Technical Compliance (Yes / No)	Additional Information (if any)
1.	TAP Reactor	I. Testing and assessment of Slide Valve function/seal.		
		II. Repair and testing of manifold function		
		III. Testing malfunctioning of pulse valve parts		
		IV. Removal and clean-up of vespel valve tip		
		V. Building valve stems		