



INDIAN INSTITUTE OF TECHNOLOGY BOMBAY

Electrical Engineering Department

Powai, Mumbai 400076.

Ref No. (PR No.1000057156)

(Rfx No.6100003117)

Technical Specifications Magnetic Shield Chamber

Sr. No	Item Description	Detailed Technical Specification	Technical Compliance (Yes / No)	Additional Information (if any)
1	Magnetic Shield Chamber	Four-layer magnetic shield set with three axis internal field coils and five first order gradient coils -Qty 01		
		Digital degaussing unit with 10 A peak output current. -Qty 01		
		Degaussing coil for MS-2. Qty 01		
		Programmable,3-channel bipolar current supply ± 10 mA output. Qty-01		
	Warranty	01 Year		